

Hepatitis C antibody test frequencies and positive rates in Switzerland from 2007 to 2017: a retrospective longitudinal study

<https://doi.org/10.57187/smw.2023.40085>

Appendix: Online Supplementary Material

Supplementary table 1: Overview of variables and calculations

| Variables | | |
|---|--|--|
| - Data source: Laboratories: | | |
| $a_{l,j}$ | Number of AB tests in laboratory l and year j | |
| $a_{l,j}^+$ | Number of positive AB tests in laboratory l and year j | |
| - Data source: FOPH (figures for all of Switzerland): | | |
| $a_{f,j}^+$ | Number of confirmed positive AB tests in year j | |
| $p_{f,j}^{a^+}$ | Number of individuals with at least one confirmed positive AB test in year j | |
| Derived quantities | | |
| Formula: | Term used in text: | Explanation: |
| $\frac{\sum_l a_{l,j}^+}{\sum_l a_{l,j}}$ | AB <i>test</i> positive rate in year j | Fraction of positive AB tests among all AB tests in year j |
| $c = \frac{a_{f,j}^+}{p_{f,j}^{a^+}}$ | Correction factor for year j | Average number of positive AB tests per positively tested individual in year j |
| $\frac{1}{c} \cdot \frac{\sum_l a_{l,j}^+}{\sum_l a_{l,j}}$ | AB <i>tested</i> positive rate in year j | Average proportion of positively tested individuals per AB test in year j |

Abbreviation: AB: antibody

Supplementary table 2: Raw data and calculated AB positive rates by year, with between- and within laboratory variability


| Description | Data source | Year | | | | | | | | | | | rsd | qcd |
|---------------------|----------------|--------------|-------|-------|-------|-------|-------|-------|-------|-------|--------|--------|--------------|-----|
| | | 2007 | 2008 | 2009 | 2010 | 2011 | 2012 | 2013 | 2014 | 2015 | 2016 | 2017 | (calculated) | |
| # AB tests | laboratory D1 | 6081 | 6619 | 6851 | 8921 | 10157 | 11680 | 13589 | 14064 | 14932 | 17427 | 17726 | 36% | 30% |
| | laboratory D2 | 3778 | 3993 | 3884 | 3862 | 3832 | 4057 | 4784 | 5296 | 5441 | 4618 | 4755 | 14% | 10% |
| | laboratory D3 | 2878 | 3929 | 4286 | 5373 | 6240 | 7631 | 8926 | 9481 | 10936 | 13049 | 14796 | 49% | 36% |
| | laboratory D4 | 6669 | 7319 | 7481 | 8230 | 8205 | 8459 | 8485 | 9680 | 8514 | 9896 | 9906 | 13% | 7% |
| | laboratory D5 | 10449 | 11121 | 11186 | 11663 | 13125 | 13656 | 15138 | 17109 | 19059 | 20664 | 21442 | 27% | 23% |
| | laboratory D6 | | | | | 1272 | 3528 | 3535 | 4136 | 8521 | 9777 | 9514 | 60% | 44% |
| | laboratory D11 | | | 1714 | 1809 | 2057 | 2397 | 2923 | 3151 | 3409 | 3605 | 3902 | 29% | 25% |
| | laboratory D12 | 234 | 2442 | 2623 | 2825 | 3326 | 3425 | 3547 | 3665 | 4236 | 5555 | 5631 | 44% | 18% |
| | laboratory D20 | 982 | 1108 | 1358 | 921 | 1114 | 1106 | 1092 | 1157 | 1368 | 1649 | 1633 | 20% | 11% |
| | laboratory F7 | 6345 | 6921 | 11199 | 12949 | 13129 | 11463 | 11960 | 10100 | 10633 | 11158 | 12557 | 21% | 8% |
| | laboratory F10 | | | 9892 | 9760 | 9822 | 10101 | 10168 | 10506 | 10847 | 11912 | 11808 | 8% | 5% |
| | laboratory I8 | 2417 | 2620 | 2938 | 2965 | 3152 | 3090 | 2777 | 3132 | 3353 | 3504 | 3332 | 11% | 6% |
| | laboratory I9 | 2272 | 2598 | 2522 | 2647 | 2710 | 2227 | 3080 | 3208 | 3577 | 4151 | 4264 | 23% | 14% |
| | total | (calculated) | 42105 | 48670 | 65934 | 71925 | 78141 | 82820 | 90004 | 94685 | 104826 | 116965 | 121266 | 31% |
| rsd | (calculated) | 74% | 63% | 67% | 69% | 73% | 67% | 68% | 67% | 64% | 65% | 66% | | |
| qcd | (calculated) | 46% | 45% | 51% | 53% | 57% | 53% | 54% | 52% | 50% | 48% | 49% | | |
| # positive AB tests | laboratory D1 | 172 | 217 | 176 | 195 | 217 | 236 | 260 | 256 | 268 | 305 | 313 | 20% | 12% |
| | laboratory D2 | 170 | 149 | 142 | 130 | 113 | 134 | 135 | 139 | 108 | 128 | 138 | 12% | 4% |
| | laboratory D3 | 117 | 129 | 168 | 186 | 183 | 213 | 256 | 233 | 217 | 265 | 278 | 26% | 16% |
| | laboratory D4 | 436 | 381 | 336 | 366 | 331 | 290 | 293 | 315 | 303 | 331 | 308 | 13% | 7% |
| | laboratory D5 | 98 | 57 | 60 | 43 | 46 | 44 | 44 | 47 | 55 | 50 | 58 | 29% | 12% |
| | laboratory D6 | | | | | 45 | 94 | 102 | 135 | 215 | 226 | 242 | 51% | 38% |
| | laboratory D11 | | | 153 | 116 | 168 | 211 | 183 | 182 | 175 | 145 | 127 | 18% | 11% |
| | laboratory D12 | 0 | 207 | 214 | 277 | 233 | 197 | 152 | 180 | 202 | 220 | 194 | 37% | 7% |
| | laboratory D20 | 44 | 40 | 33 | 30 | 31 | 34 | 45 | 43 | 54 | 62 | 59 | 26% | 19% |
| | laboratory F7 | 67 | 65 | 125 | 183 | 154 | 139 | 125 | 104 | 106 | 113 | 99 | 30% | 13% |
| | laboratory F10 | | | 311 | 258 | 294 | 267 | 273 | 253 | 299 | 282 | 300 | 7% | 6% |
| | laboratory I8 | 117 | 94 | 134 | 106 | 130 | 93 | 94 | 107 | 77 | 80 | 82 | 19% | 12% |
| | laboratory I9 | 139 | 150 | 149 | 149 | 145 | 132 | 146 | 146 | 147 | 148 | 181 | 8% | 1% |
| | total | (calculated) | 1360 | 1489 | 2001 | 2039 | 2090 | 2084 | 2108 | 2140 | 2226 | 2355 | 2379 | 16% |
| rsd | (calculated) | 87% | 68% | 53% | 57% | 58% | 51% | 52% | 50% | 51% | 53% | 53% | | |
| qcd | (calculated) | 37% | 45% | 17% | 30% | 32% | 39% | 43% | 37% | 34% | 40% | 47% | | |


| | | | | | | | | | | | | | | | |
|--|------------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|-------|-----|-----|
| <i>AB test positive rate</i> | (calculated D1) | 0.028 | 0.033 | 0.026 | 0.022 | 0.021 | 0.020 | 0.019 | 0.018 | 0.018 | 0.018 | 0.018 | 23% | 14% | |
| | (calculated D2) | 0.045 | 0.037 | 0.037 | 0.034 | 0.029 | 0.033 | 0.028 | 0.026 | 0.020 | 0.028 | 0.029 | 21% | 11% | |
| | (calculated D3) | 0.041 | 0.033 | 0.039 | 0.035 | 0.029 | 0.028 | 0.029 | 0.025 | 0.020 | 0.020 | 0.019 | 26% | 20% | |
| | (calculated D4) | 0.065 | 0.052 | 0.045 | 0.044 | 0.040 | 0.034 | 0.035 | 0.033 | 0.036 | 0.033 | 0.031 | 26% | 14% | |
| | (calculated D5) | 0.009 | 0.005 | 0.005 | 0.004 | 0.004 | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | 0.002 | 0.003 | 51% | 22% |
| | (calculated D6) | | | | | 0.035 | 0.027 | 0.029 | 0.033 | 0.025 | 0.023 | 0.025 | 16% | 10% | |
| | (calculated D11) | | | 0.089 | 0.064 | 0.082 | 0.088 | 0.063 | 0.058 | 0.051 | 0.040 | 0.033 | 32% | 23% | |
| | (calculated D12) | 0.000 | 0.085 | 0.082 | 0.098 | 0.070 | 0.058 | 0.043 | 0.049 | 0.048 | 0.040 | 0.034 | 50% | 30% | |
| | (calculated D20) | 0.045 | 0.036 | 0.024 | 0.033 | 0.028 | 0.031 | 0.041 | 0.037 | 0.039 | 0.038 | 0.036 | 17% | 10% | |
| | (calculated F7) | 0.011 | 0.009 | 0.011 | 0.014 | 0.012 | 0.012 | 0.010 | 0.010 | 0.010 | 0.010 | 0.008 | 15% | 7% | |
| | (calculated F10) | | | 0.031 | 0.026 | 0.030 | 0.026 | 0.027 | 0.024 | 0.028 | 0.024 | 0.025 | 9% | 4% | |
| | (calculated I8) | 0.048 | 0.036 | 0.046 | 0.036 | 0.041 | 0.030 | 0.034 | 0.034 | 0.023 | 0.023 | 0.025 | 25% | 17% | |
| | (calculated I9) | 0.061 | 0.058 | 0.059 | 0.056 | 0.054 | 0.059 | 0.047 | 0.046 | 0.041 | 0.036 | 0.042 | 17% | 14% | |
| overall | (calculated) | 0.032 | 0.031 | 0.030 | 0.028 | 0.027 | 0.025 | 0.023 | 0.023 | 0.021 | 0.020 | 0.020 | 18% | 14% | |
| rsd | (calculated) | 63% | 60% | 62% | 64% | 59% | 64% | 50% | 50% | 52% | 45% | 44% | | | |
| qcd | (calculated) | 52% | 19% | 32% | 30% | 19% | 13% | 21% | 21% | 33% | 27% | 27% | | | |
| # positive (conf.) AB tests | FOPH | | | | | | 2065 | 1959 | 2057 | 1959 | 2129 | 2002 | 3% | 2% | |
| # individuals with ≥ 1 (conf.) positive AB test | FOPH | | | | | | 1797 | 1694 | 1743 | 1745 | 1840 | 1782 | 3% | 1% | |
| Correction factor | (calculated) | | | | | | 1.149 | 1.156 | 1.180 | 1.123 | 1.157 | 1.123 | 2% | 1% | |
| <i>AB tested positive rate</i> | (calculated D1) | | | | | | 0.018 | 0.017 | 0.015 | 0.016 | 0.015 | 0.016 | 6% | 3% | |
| | (calculated D2) | | | | | | 0.029 | 0.024 | 0.022 | 0.018 | 0.024 | 0.026 | 16% | 6% | |
| | (calculated D3) | | | | | | 0.024 | 0.025 | 0.021 | 0.018 | 0.018 | 0.017 | 18% | 14% | |
| | (calculated D4) | | | | | | 0.030 | 0.030 | 0.028 | 0.032 | 0.029 | 0.028 | 5% | 3% | |
| | (calculated D5) | | | | | | 0.003 | 0.003 | 0.002 | 0.003 | 0.002 | 0.002 | 10% | 4% | |
| | (calculated D6) | | | | | | 0.023 | 0.025 | 0.028 | 0.022 | 0.020 | 0.023 | 11% | 4% | |
| | (calculated D11) | | | | | | 0.077 | 0.054 | 0.049 | 0.046 | 0.035 | 0.029 | 35% | 17% | |
| | (calculated D12) | | | | | | 0.050 | 0.037 | 0.042 | 0.042 | 0.034 | 0.031 | 17% | 9% | |
| | (calculated D20) | | | | | | 0.027 | 0.036 | 0.031 | 0.035 | 0.032 | 0.032 | 10% | 4% | |
| | (calculated F7) | | | | | | 0.011 | 0.009 | 0.009 | 0.009 | 0.009 | 0.007 | 13% | 2% | |
| | (calculated F10) | | | | | | 0.023 | 0.023 | 0.020 | 0.025 | 0.020 | 0.023 | 7% | 5% | |
| | (calculated I8) | | | | | | 0.026 | 0.029 | 0.029 | 0.020 | 0.020 | 0.022 | 17% | 15% | |
| | (calculated I9) | | | | | | 0.052 | 0.041 | 0.039 | 0.037 | 0.031 | 0.038 | 17% | 5% | |
| overall | (calculated) | | | | | | 0.022 | 0.020 | 0.019 | 0.019 | 0.017 | 0.017 | 9% | 6% | |
| rsd | (calculated) | | | | | | 64% | 50% | 50% | 52% | 45% | 44% | | | |
| qcd | (calculated) | | | | | | 13% | 21% | 21% | 33% | 27% | 27% | | | |

Abbreviations: AB: antibody; rsd: relative standard deviation (= coefficient of variance); qcd: quartile coefficient of dispersion

Supplementary methods 1: Exploration of the between-laboratory variability

As mentioned above, we assume that our data covers approximately 80% of all HCV AB tests in Switzerland. Since the data stem from different-sized laboratories in different parts of the country, the variability in numbers and rates between laboratories (and - for comparison - within the same laboratory over time) may be of interest. In order to explore this variability, we present individual values per laboratory and year of the number of AB tests (A.), the number of positive AB tests (B.), the AB *test* positive rate (C.) and the AB *tested* positive rate (D.) and, additionally, calculated their between-laboratory variabilities (E.) using both a non-robust and a robust measure (Supplementary table 2). For better illustration, counts and rates and the variability of the AB *test* positive rate over time are shown graphically in the following Supplementary figure x, facets A.-D. and E..





E.

Supplementary figure 1: A. Number of HCV AB tests, B. number of positive HCV AB tests, C. AB *test* positive rate, D. AB *tested* positive rate, by laboratory and year, and E. between-laboratory variability the AB *test* positive rate per year. (Between-laboratory variability measures of the AB *tested* positive rates are not shown, since they coincide, where defined, with the corresponding variability measures of the AB *test* positive rates.) **Abbreviations:** AB: antibody; rsd: relative standard deviation (= coefficient of variance); qcd: quartile coefficient of dispersion.

The relative standard deviation (rsd) as the less robust measure of the between-laboratory lies within a range from 44% to 63%, and the more robust quartile coefficient of dispersion (qcd) ranges from 13% to 52%. The rsd shows a tendency to decrease in the more recent years which can be explained mainly by the rates in laboratories D11 and D12 converging towards the mean rates. The relatively stable qcd confirms that overall, AB *test* positive rates (and thus AB *tested* positive rates) remain on a moderate level, and that the decrease in between-laboratory variability can be attributed to few outliers who contributed few tests before 2012. Overall, the between-laboratory variabilities are about twice as large as the within-laboratory variabilities over time (rsd: min 9% - max 51%; qcd: min 4% - max 30%).

Supplementary methods 2: Representativeness of the study sample

We assume that the data from our sample of 13 laboratories is reasonably representative for all HCV antibody testing in Switzerland for the following reasons:

- First, according to the FAMH (the medical laboratories of Switzerland), there are only 43 laboratories that potentially determine HCV antibodies, and our sample contains 13 of the 20 largest of these. Notably, all three major language regions of Switzerland are represented among them.
- Secondly, and even more importantly, if the total numbers of positive HCV AB tests provided by the laboratories in our sample is tentatively multiplied with a positive predictive value (PPV) of 72% (as suggested by Cadieux et al. 2016; see below) then this yields estimated numbers of confirmed tests close to the numbers provided by the Federal Office of Public Health (FOPH) for *all* Switzerland. With the numbers already provided in Supplementary table 2:

| Year | 2012 | 2013 | 2014 | 2015 | 2016 | 2017 |
|---|------|------|------|------|------|------|
| # positive AB tests (13 laboratories): | 2084 | 2108 | 2140 | 2226 | 2355 | 2379 |
| estimated # confirmed positive AB tests (13 laboratories) = # positive AB tests (13 laboratories) * 0.72 | 1500 | 1518 | 1541 | 1603 | 1696 | 1713 |
| # positive confirmed AB tests (all Switzerland, according to FOPH): | 2065 | 1959 | 2057 | 1959 | 2129 | 2002 |
| estimated fraction of confirmed tests covered by study sample, by year | 73% | 77% | 75% | 82% | 80% | 86% |
| estimated fraction of confirmed tests covered by study sample, mean over all years | 79% | | | | | |

Abbreviations: AB: antibody; FOPH: Federal Office of Public Health

Thus, our study includes roughly 73%-86% of all positive HCV AB tests from all of Switzerland in 2012-2017, suggesting a reasonable degree of representativeness of our laboratories and our data.

Reference: Cadieux G, Campbell J, Dendukuri N. Systematic review of the accuracy of antibody tests used to screen asymptomatic adults for hepatitis C infection. *CMAJ Open*. 2016;4(4):E737-E745. doi:10.9778/cmajo.20160084